





Linearity of sensor system (uncalibrated)




Deviation from reference (uncalibrated)




Average linear regression coefficients (uncalibrated)

$y = ax+b$	a	b	n
Alphasense	0.81	4	5
Cairclip	0.74	0	3
Membrapor C1	0.48	13	5
Membrapor C2	0.88	9	2


Sensor stability (calibrated)



Between-sensor uncertainty (calibrated)




Sensor response to change in relative humidity





Uncalibrated sensor versus reference




Accuracy (uncalibrated)

Reference mean (µg/m ³)	Sensor mean (µg/m ³)	Accuracy (%)
72	62	85
110	96	87
146	123	84
181	156	86
216	178	82
267	217	81

average
84/100

Linearity before calibration



- $r^2 > 0.99$
- slope ≈ 0.8 ;
- intercept (µg/m³): $\ll -5 \text{ --- } +16 \gg$
(used for calibration)

Steady-state stability

$< 2 \mu\text{g}/\text{m}^3$
(calibrated)

Between sensor uncertainty

$6.5 \mu\text{g}/\text{m}^3$
(calibrated)


Influence of T and RH

large
(T: 5-35°C; RH: 45-90%)

Ozone interference

small
(O₃: 30-160 ppb)

Uncalibrated sensor versus reference




Accuracy (uncalibrated)

Reference mean (µg/m ³)	Sensor mean (µg/m ³)	Accuracy (%)
72	50	69
110	82	75
146	110	75
181	131	72
216	164	76
267	199	75

average
74/100

Linearity before calibration



- $r^2 > 0.99$
- slope ≈ 0.74 ;
- intercept (µg/m³): $\ll -8 \text{ --- } 11 \gg$
(used for calibration)

Steady-state stability

$< 4 \mu\text{g}/\text{m}^3$
(calibrated)

Between sensor uncertainty

$11.5 \mu\text{g}/\text{m}^3$
(calibrated)


Influence T/RH

moderate
(T: 5-35°C; RH: 45-90%)

Ozone interference

moderate
(O₃: 30-160 ppb)

Uncalibrated sensor versus reference




Accuracy (uncalibrated)

Reference mean (µg/m ³)	Sensor mean (µg/m ³)	Accuracy (%)
72	46	64
110	65	60
146	84	57
181	100	55
216	113	52
267	137	51

average
57/100

Linearity before calibration



- $r^2 > 0.99$
- slope: << 0.14 --- 0.6 >>
- intercept (µg/m³): << 5 --- 28 >>
(used for calibration)

Steady-state stability

< 3 µg/m³
(calibrated)

Between sensor uncertainty

8.2 µg/m³
(calibrated)

Influence T/RH


moderate

Ozone interference

large

(O₃: 30-160 ppb)

Uncalibrated sensor versus reference




Accuracy (uncalibrated data)

Reference mean (ug/m3)	Sensor mean (ug/m3)	Accuracy (%)
72	71	97
110	109	100
146	138	94
181	163	90
216	201	93
267	240	90

average
94/100

Linearity before calibration



- $r^2 > 0.99$
- slope ≈ 0.88 ;
- intercept ($\mu\text{g}/\text{m}^3$): $\ll 7 \text{ --- } 10 \gg$
(used for calibration)

Steady-state stability

$< 4 \mu\text{g}/\text{m}^3$
(calibrated)

Between sensor uncertainty

$9.9 \mu\text{g}/\text{m}^3$
(calibrated)

Influence T/RH

moderate
(T: 5-35°C; RH: 45-90%)

Ozone interference

large
(O₃: 30-160 ppb)